

TSS 500N10

TELECOM SURGE GENERATOR AS PER ITU / FCC UP TO 10 KV



FOR TESTS ACCORDING TO ...

- > EN 60950-1
- > EN 61000-4-5
- > FCC 97-270 (part 68)
- > IEC 60065
- > IEC 60950-1
- > IEC 61000-4-5
- > ITU-T K.17
- > ITU-T K.20
- > ITU-T K.21
- > ITU-T K.28
- > ITU-T K.45

TSS 500N10 - 10 KV TELECOM SURGE GENERATOR FOR 10/700 US TEST PULSE







Telecommunication networks are exposed to lightning events. Therefore telecommunication equipment being connected to the outside world need to have appropriate protection to show an acceptable immunity to surge transients in order not to fail in case of lightning events. Telecom Surge Simulators of the TSS 500N series are used to proof the immunity of telecommunications equipment.

The TSS 500N10 is used to perform tests as per EN/IEC 61000-4-5 and related standards and complies with the requirements of ITU-T and FCC 97-270 (part 68) for Surge B pulse. It also generates the Surge voltage pulse 1.2/50 us as per EN/IEC 60950.

HIGHLIGHTS

- > **Standalone tester for 10/700 us pulse as per EN/IEC 61000-4-5**
- > **Test voltage up to 10 kV**
- > **Complies to ITU-T for enhanced level testing**
- > **Complies to FCC part 68 (Surge B pulses)**
- > **Built-in CDNs for 2-wire and 4-wire applications**

APPLICATION AREAS

- | | |
|--|---|
|  INDUSTRY |  TELECOM |
|  COMPONENTS |  RESIDENTIAL |
|  MEDICAL | |
|  BROADCAST | |

TECHNICAL DETAILS

TELECOM SURGE GENERATOR

AS PER ITU AND ETS RECOMMENDATIONS	
	Pulse 1.2/50 us
Voltage (o.c.)	500 V - 10,000 V ± 10%
Rise time*)	1 us ± 30%
Pulse duration*)	50 us ± 20%
Energy storage capacitor	1 uF
	Pulse 10/700 us
Voltage (o.c.)	500 V - 10,000 V ± 10%
Rise time*)	6.5 us ± 30%
Pulse duration*)	700 us ± 20%
Energy storage capacitor	20 uF
Polarity	Positive, negative or alternating
Counter	1 - 30,000 or endless
*) definition of waveform parameters as per IEC 60469-1. As per IEC 61000-4-5 this is considered to be equal to the waveform parameter definition as per IEC 60060-1 for the 1.2/50us pulse and CCITT for the 10/700us pulse.	

AS PER FCC PART 68	
	Surge Pulse B
Voltage (o.c.)	500 V - 10,000 V ± 10%
Front time	9 us ± 30%
Pulse duration	720 us ± 20%
Current (s.c.)	12.5 - 250 A for T1 to COM or T2 to COM
Rise time	5 us ± 30%
Pulse duration	320 us ± 20%
Energy storage capacitor	20 uF
Polarity	Positive, negative or alternating
Counter	1 - 30,000 or endless

TELECOM SURGE GENERATOR

AS PER EN/IEC 61000-4-5	
	Pulse 10/700 s
Voltage (o.c.)	500 V - 10,000 V ± 10%
Rise time*)	10 us ± 30%
Pulse duration*)	700 us ± 20%
Current (s.c.)	12.5 - 250 A for T1 to COM or T2 to COM
Rise time*)	5 us ± 20%
Pulse duration*)	320 us ± 20%
Energy storage capacitor	20 uF
Source impedance	40 ohm
Polarity	Positive, negative or alternating
Counter	1 - 30,000 or endless

COUPLING AS PER	
ITU-T	2-wire: T1 and T2 with 25 ohm each
FCC part 68	2-wire: T1 and T2 with 25 ohm each
EN/IEC 61000-4-5	4-wire: T1, T2, T3, T4 with 25 ohm each

TRIGGER	
Automatic	Automatic pulse release
Manual	Single pulse release
External	External pulse release
CRO trigger	5V trigger signal for oscilloscope

TECHNICAL DETAILS

GENERAL DATA

TEST ROUTINES

Quick Start	Immediate start; easy-to-use and fast
User Test routines	Change Polarity after n pulses Change voltage after n pulses
Standard Test routines	IEC 61000-4-5 Level 1,000 V IEC 61000-4-5 Level 2,000 V IEC 61000-4-5 Level 4,000 V FCC part 68, Pulse B Metallic 1,000 V FCC part 68, Pulse B Longitudinal 1,500 V
Service	Service, setup, self test

INTERFACE

Optical interface	Opto link, 3 m cable USB A connector
Parallel interface	IEEE 488, addresses 1 - 30

SAFETY

Safety circuit	Control input (24V dc)
Warning lamp	Floating output contact

GENERAL DATA

Dimensions, weight	19"/9 HU, 420 mm x 448 mm x 500 mm, 37.1 kg
Supply voltage	115/230 V +10/-15%
Fuses	2x 2 AT (230V) or 2x 4 AT (115V)
Temperature	10° C to 35° C
Rel. humidity	Max. 85 %, non condensing
Atmospheric pressure	86 kPa (860 mbar) to 106 kPa (1,060 mbar)

ALTERNATIVE CONFIGURATIONS

T1-N10	0.5/700us up to 10kV
T2-N10	100/700us up to 10kV
	to replace one of the standard pulses

OPTIONS

OPTIONS

CNV 504T5	Coupling/decoupling network for unshielded symmetrical lines (communication lines) as per IEC/EN 61000-4-5 Ed.3 (fig. 10) for 4 lines.
CNV 508T8	Coupling/decoupling network for unshielded symmetrical lines (communication lines) as per IEC/EN 61000-4-5 Ed.3 (fig. 10) for 8 lines.
iec.control 1	Software to control the test, including standard library, test report facility and data conversion generator

COMPETENCE WHEREVER YOU ARE



CONTACT EM TEST DIRECTLY

Switzerland

AMETEK CTS GmbH › Sternenhofstraße 15 › 4153 Reinach › Switzerland
Phone +41 (0)61 204 41 11 › Fax +41 (0)61 204 41 00
Internet: www.ametek-cts.com › E-mail: sales.conducted.cts@ametek.com

Germany

AMETEK CTS Europe GmbH › Customer Care Center EMEA › Lünener Straße 211
› 59174 Kamen › Germany
Phone +49 (0) 2307 26070-0 › Fax +49 (0) 2307 17050
Internet: www.ametek-cts.com › E-mail: info.cts.de@ametek.com

Poland

AMETEK CTS Europe GmbH › Biuro w Polsce › ul. Twarda 44 › 00-831 Warsaw › Poland
Phone +48 (0) 518 643 12
Internet: www.ametek-cts.com › E-mail: Infopolska.cts@ametek.com

USA / Canada

AMETEK CTS US › 52 Mayfield Ave › Edison › NJ 08837 › USA
Phone +1 732 417 0501
Internet: www.ametek-cts.com › E-mail: usasales.cts@ametek.com

P.R. China

AMETEK Commercial Enterprise (Shanghai) Co. Ltd. › Beijing Branch › Western Section, 2nd floor › Jing Dong Fang Building (B10) › Chaoyang District › Beijing, China, 100015
Phone +86 10 8526 2111 › Fax +86 (0)10 82 67 62 38
Internet: www.ametek-cts.com › E-mail: chinasales@ametek.com

Republic of Korea

EM TEST Korea Limited › #405 › WooYeon Plaza › #986-8 › YoungDeok-dong › Giheung-gu › Yongin-si › Gyeonggi-do › Korea
Phone +82 (31) 216 8616 › Fax +82 (31) 216 8616
Internet: www.emtest.co.kr › E-mail: sales@emtest.co.kr

Singapore

AMETEK Singapore Pte. Ltd › No. 43 Changi South Avenue 2 › 04-01 Singapore 48164
Internet: www.ametek-cts.com › E-mail: singaporesales.cts@ametek.com

Great Britain

AMETEK GB › 5 Ashville Way › Molly Millars Lane › Wokingham › Berkshire RG41 2 PL › Great Britain
Phone +44 845 074 0660
Internet: www.ametek-cts.com

Information about scope of delivery, visual design and technical data correspond with the state of development at time of release. Subject to change without further notice.